

**Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 3/9/2015**

*LMS
JPD*

Instrument Serial Number: 010619

As of 09-Mar-15

Start Date	End Date	Agency Name
04-Mar-15		DFS Central Lab
23-Sep-14	04-Mar-15	Manassas Park PD
10-Sep-14	23-Sep-14	DFS Central Lab
09-Apr-14	10-Sep-14	Floyd County Sheriff's Office
14-Jan-14	09-Apr-14	DFS Central Lab
28-Feb-13	14-Jan-14	Fairfax County PD
03-Sep-10	28-Feb-13	DFS Central Lab
21-Apr-09	03-Sep-10	Fairfax County PD
04-Aug-08	21-Apr-09	DFS Central Lab

HMS
JPD

INTOX EC/IR II

Quality Assurance Worksheet

Instrument Serial Number **010619** Worksheet Start Date **3/4/2015**

Location **Manassas Park Police Department**

Address **329 Manassas Dr., Manassas Park, VA 20111**

DFS Technician **Heather Stanton** License No. **18910**

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) **753** Reference Barometer (mm HG) **755**

Reference Barometer(RB)Serial # **009113** RB Calibration Due **4/28/2015**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	
0.300	0.297	0.288	0.305	0.299	
Precision		sample min	sample max	Sample 2	
0.001		0.299	0.300	0.300	
				Sample 3	

Standard (sea level)	PA Target	minimum	maximum	Sample 1	
0.100	0.099	0.096	0.102	0.098	
Precision		sample min	sample max	Sample 2	
0		0.098	0.098	0.098	
				Sample 3	


Dry gas standard Lot No. (with tank no.) **AG320502-12**

☐ Replaced dry gas standard (+O-ring)


☐ Installed at Location

☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:



Instrument Serial Number

010619

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath M. Stark

Date

3/9/15

Issuing Analyst

Jenne R. Fly

Date

3/9/15

MSK




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

HMS
JPD

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 03/04/2015
TEST EQUIPMENT NUMBER 010619		

RESULTS: TIME SAMPLE TAKEN 10:42 EST

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 04 Mar 2015 at 11:23

AMS
JPD

Test Results

Instrument Serial Number 010619

Test # 002666 Subject Test

Test Location 1 Department of
Test Date 04 Mar 2015

Test Location 2 Forensic Science
Test Time 10:36
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial N

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 04 Mar 2015 End Time 10:43

Result Time 10:42

Result Date 04 Mar 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:36

Data Type BLK

Sample Value 0.000

Sample Time 10:37

Data Type CHK

Sample Value 0.098

Sample Time 10:38

Data Type BLK

Sample Value 0.000

Sample Time 10:39

Data Type SUBJ

Sample Value 0.000

Sample Time 10:40

Data Type BLK

Sample Value 0.000

Sample Time 10:41

Data Type SUBJ

Sample Value 0.000

Sample Time 10:42

Data Type BLK

Sample Value 0.000

Sample Time 10:43

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG320502-12

Standard Expiration Date 07/24/2015

Tank Pressure 257

Barometric Pressure 753 mmHg

Blow Sample Number 1 Blow Duration 3.19 sec

Blow Volume 1753 cc End-of-Blow Time 10:40

Blow Sample Number 2 Blow Duration 3.19 sec

Blow Volume 1740 cc End-of-Blow Time 10:42

Tamper Evident Stamp e051395d

Test Status Code 0

Test Status Success

HMS
JPD

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010619 Test Number: 2667

Test Date: 03/04/2015 Test Time: 10:44 EST

Dry Gas Target: 0.099

Lot Number: AG320502-12 Exp Date: 07/24/2015

Tank Pressure: 252 psi Barometric Pressure: 752 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:45
CHK	0.098	10:46
BLK	0.000	10:47
CHK	0.098	10:48
BLK	0.000	10:49
CHK	0.098	10:50

Test Status: *Success*

Calibration CRC: 92BB2488

the
JF12

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010619 Test Number: 2668

Test Date: 03/04/2015 Test Time: 10:53 EST

Dry Gas Target: 0.297

Lot Number: AG428003-07 Exp Date: 10/07/2016

Tank Pressure: 90 psi Barometric Pressure: 752 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:54
CHK	0.299	10:55
BLK	0.000	10:56
CHK	0.300	10:57
BLK	0.000	10:59
CHK	0.300	11:00

Test Status: *Success*

Calibration CRC: 92BB2488